

- Industry 4.0 adopted sub-micron Computerized test indicator with computerized MICS system color Touchscreen
- Patents UA126928, UA99686
- Functions: Go/NoGo, Max/Min, Formula, Timer, Linear correction, Temperature correction, Resolution selecting, Connection status
- Memory manager: 2000 values, Folders system, Statistics mode, Memory data transfer
- Quick reading system (50 value/sec)
- **4 MODES DATA TRANSFER:** WIRELESS to MDS app (Windows, Android, iOS, MacOS); WIRELESS HID, WIRELESS HID+MAC, USB HID
- Calibration certificate included (ISO17025 (Ilac MRA))



WIRELESS TO MDS APP  
WIRELESS HID+MAC  
USB HID

Item No	Range		Resol.	Accu- racy	Probe	Measu- ring Force	Preset	Go/NoGo	Max/Min	Formula	Timer	Temp comp	Linear corr	Calibr date	FW update	Recharge	Memory	WIRELESS	USB	
	mm	inch	mm	μm	mm	N														
120129018	-0.8 +0.8	-0.03" +0.03"	<b>0,0001</b>	±5	12,8	0,1-0,18	•	•	•	•	•	•	•	•	•	•	•	•	•	•

## SUB-MICRON COMPUTERIZED TEST INDICATOR EXTERNAL PROBE



WIRELESS TO MDS APP  
WIRELESS HID+MAC  
USB HID

Item No	Range		Resol.	Accu- racy	Probe	Measu- ring Force	Cable length	Preset	Go/NoGo	Max/Min	Formula	Timer	Temp comp	Linear corr	Calibr date	FW update	Recharge	Memory	WIRELESS	USB	
	mm	inch	mm	μm	mm	N	m														
120129019L	-0.8	-0.03"	<b>0,0001</b>	±5	12,8	0,1-0,18	2	•	•	•	•	•	•	•	•	•	•	•	•	•	•
120129019L10	+0.8	+0.03"			12,8		10	•	•	•	•	•	•	•	•	•	•	•	•	•	•

DOWNLOAD APP

